

Declaration of Compliance

The M400 and M4000 were tested and evaluated to the following standards: IEC 60601-1:2005, AMD1:2012

IEC 60601-1-6:2010, AMD1:2013

IEC 60601-1-11:2015

IEC 60601-1-2:2014

The following national deviations were included in the evaluation:

Country	Standard	Differences*
Albania	SSH EN 60601-1:2006 + A1:2013 + AC:2014	<input type="checkbox"/>
Austria (AU)	ÖVE/ÖNORM EN 60601-1: 2014 02 01	<input type="checkbox"/>
Brazil (BR)	NBR IEC 60601-1:2010 + IEC 60601-1:2005/AMD1:2012	<input type="checkbox"/>
Bosna Hercegovina	BAS EN 60601-1:2010 + A1:2015 + A12:2016	<input type="checkbox"/>
Bulgaria	БДC EN 60601-1:2006 + A1:2013 + AC:2014	<input type="checkbox"/>
Canada (CA)	CAN/CSA C22.2 No. 60601-1:2014	<input checked="" type="checkbox"/>
Denmark (DK)	DS/EN 60601-1:2006 + A1:2013 + AC:2014	<input type="checkbox"/>
European Union (EU)	EN 60601-1:2006 + A1:2013 + A12:2014 + AC:2014	<input type="checkbox"/>
Finland (FI)	SFS-EN 60601-1:2007 + A1:2016	<input type="checkbox"/>
Former Yugoslav Republic of Macedonia	MCK EN 60601-1:2006 + A1:2016 + AC:2015	<input type="checkbox"/>
France (FR)	NF EN 60601-1:2007 + A1:2014	<input type="checkbox"/>
Germany (DE)	DIN EN 60601-1:2013; VDE 0750-1:2013-12	<input type="checkbox"/>
Hungary (HU)	MSZ EN 60601-1:2017	<input type="checkbox"/>
Iceland	IST EN 60601-1:2006 + A1:2013 + AC:2014	<input type="checkbox"/>
Ireland	I.S. EN 60601-1:2006 + AMD1:2013 + AMD12:2014	<input type="checkbox"/>
Israel	SI 60601-1 Part 1	<input checked="" type="checkbox"/>
Korea	IEC 60601-1:2012 / MDFS No. 2020-12, Annex 1	<input checked="" type="checkbox"/>
Latvia	LVS EN 60601-1:2007 + A1:2014	<input type="checkbox"/>
Lithuania	LST EN 60601-1:2007 + A1:2013	<input type="checkbox"/>
Luxembourg	ILNAS-EN 60601-1:2006 + A1:2014	<input type="checkbox"/>
Montenegro	MEST EN 60601-1:2010 + A1:2015 + A12:2015	<input type="checkbox"/>
Netherlands (NO)	NEN-EN-IEC 60601-1:2006 + A1:2013 + A12:2014	<input type="checkbox"/>
Norway (NO)	NEK EN 60601-1:2006 + A1:2013	<input type="checkbox"/>
Poland (PL)	PN-EN 60601-1:2011 + A12:2014 AC1:2015	<input type="checkbox"/>
Serbia	SPRS EN 60601-1:2012 + A1:2014 + AC:2017	<input type="checkbox"/>
Slovenia (SL)	SIST EN 60601-1:2007 + A1:2014 + AC:2015	<input type="checkbox"/>
Spain	UNE-EN 60601-1:2008 + A1:2013 + A12:2014	<input type="checkbox"/>
Sweden (SE)	SS-EN 60601-1:2006 + A1:2013 + A11:2014	<input type="checkbox"/>

Switzerland (CH)	SN EN 60601-1:2010 + A1:2013 + Corr:2014	<input type="checkbox"/>
Turkey (TR)	TS EN 60601-1:2007 + A12:2015 + AC:2017 + A12:2017 + A1:2017	<input type="checkbox"/>
United States of America (US)	ANSI/AAMI ES60601-1:2005/(R)2012 and A1:2012, C1:2009/(R)2012 and A2:2010/(R)2012	<input checked="" type="checkbox"/>
United Kingdom (UK)	BS EN 60601-1:2006 + A12:2014	<input type="checkbox"/>
<i>*Standards with identified differences are covered; standards with no differences are considered covered based on compliance with IEC 60601-1:2005 + A1:2012. Identification of differences is based either on the referenced National/Regional Standard or on the IECCEB CB Scheme list of declared differences. NOTE: where the publisher of a National/Regional Standard identifies their standard as identical (idt) or equivalent (eqv) it is assumed to be identical to IEC 60601-1:2005 + A1:2012</i>		

This evaluation has been completed by MECA, LLC for Vuzix Corporation and the M400 and M4000 were found to be in compliance with the applicable standards listed above.

Reference Report 20201167-1-0 for full compliance information.

472RP0002 (0001) - M400_M4000 IEC60601-1 Test Report



EMI Test Report 2010-046E Rev. N/C

Test Standards: IEC 60601-1-2:2014 (4th Edition)

For

Vuzix Corporation

25 Hendrix Road, Suite A
West Henrietta, NY 14586

On

#472 Vuzix M400 Smart Glasses

Model Number: N/A; Part Number: N/A; Serial Number: N/A; Software/Firmware Edition: N/A

Performed By: **Keystone Compliance, LLC.**
131 Columbus Inner Belt
New Castle, PA 16101


Keystone Compliance, LLC. does hereby certify that all inspections and tests have been performed in accordance with the documents referenced herein with exceptions as noted in this report. The results in this report pertain to the specified equipment tested, as received. This report shall not be reproduced, except in full, without the written authorization of Keystone Compliance, LLC.

Prepared By:


Maria Rodgers, Technical Writer


Date: 12/16/2020

Approved By:


Tony Masone Jr., EMC Lab Manager

Date: 12/16/2020

Approved By:


Joey Sullivan, Quality Manager

Date: 12/16/2020

Title page only - Test Report Available

472RP0001 (0001) M400 IEC60601-1-2:2014 EMI Test Report



EMI Test Report 2011-124E Rev. N/C

Test Standards: IEC 60601-1-2:2014 (4th Edition)

For

Vuzix Corporation

25 Hendrix Road, Suite A
West Henrietta, NY 14586

On

Vuzix M4000 Smart Glasses

Model Number: 490; Part Number: N/A; Serial Number: N/A; Software/Firmware Edition: 1.0

Performed By: **Keystone Compliance, LLC.**
131 Columbus Inner Belt
New Castle, PA 16101


Keystone Compliance, LLC. does hereby certify that all inspections and tests have been performed in accordance with the documents referenced herein with exceptions as noted in this report. The results in this report pertain to the specified equipment tested, as received. This report shall not be reproduced, except in full, without the written authorization of Keystone Compliance, LLC.

Prepared By:


Maria Rodgers, Technical Writer


Date: 1/22/2021

Approved By:


Tony Masone Jr., EMC Lab Manager

Date: 1/22/2021

Approved By:


Joey Sullivan, Quality Manager

Date: 1/22/2021

title page only - Test Report Available

490RP0001 (0001) M4000 IEC60601-1-2:2014 EMI Test Report